

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092923\
 Data File : VX037989.D
 Acq On : 29 Sep 2023 20:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 29 23:11:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.527	0.413	21.6	69	0.00
3 P	Chloromethane	0.573	0.496	13.4	81	0.00
4 C	Vinyl Chloride	0.619	0.565	8.7#	80	0.00
5 T	Bromomethane	0.443	0.436	1.6	84	0.00
6 T	Chloroethane	0.426	0.401	5.9	85	0.00
7 T	Trichlorofluoromethane	0.995	0.996	-0.1	86	0.00
8 T	Diethyl Ether	0.385	0.376	2.3	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.505	0.511	-1.2	88	0.00
10 T	Methyl Iodide	0.668	0.680	-1.8	85	0.00
11 T	Tert butyl alcohol	0.141	0.120	14.9	76	0.00
12 CM	1,1-Dichloroethene	0.488	0.498	-2.0#	88	0.00
13 T	Acrolein	0.139	0.093	33.1#	59	0.00
14 T	Allyl chloride	0.812	0.767	5.5	83	0.00
15 T	Acrylonitrile	0.314	0.299	4.8	83	0.00
16 T	Acetone	0.316	0.237	25.0#	71	0.00
17 T	Carbon Disulfide	1.223	1.209	1.1	86	0.00
18 T	Methyl Acetate	1.140	1.056	7.4	82	0.00
19 T	Methyl tert-butyl Ether	1.794	1.781	0.7	88	-0.01
20 T	Methylene Chloride	0.641	0.605	5.6	90	0.00
21 T	trans-1,2-Dichloroethene	0.542	0.580	-7.0	93	0.00
22 T	Diisopropyl ether	1.692	1.682	0.6	86	0.00
23 T	Vinyl Acetate	1.245	1.231	1.1	84	0.00
24 P	1,1-Dichloroethane	1.014	1.007	0.7	88	0.00
25 T	2-Butanone	0.446	0.388	13.0	77	-0.01
26 T	2,2-Dichloropropane	0.819	0.734	10.4	79	0.00
27 T	cis-1,2-Dichloroethene	0.664	0.689	-3.8	90	0.00
28 T	Bromochloromethane	0.473	0.402	15.0	77	0.00
29 T	Tetrahydrofuran	0.277	0.252	9.0	79	-0.01
30 C	Chloroform	1.092	1.118	-2.4#	90	0.00
31 T	Cyclohexane	0.843	0.855	-1.4	86	0.00
32 T	1,1,1-Trichloroethane	0.925	0.979	-5.8	89	0.00
33 S	1,2-Dichloroethane-d4	0.657	0.654	0.5	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.335	0.358	-6.9	91	0.00
36 T	1,1-Dichloropropene	0.461	0.485	-5.2	90	0.00
37 T	Ethyl Acetate	0.507	0.479	5.5	81	-0.02
38 T	Carbon Tetrachloride	0.482	0.520	-7.9	90	0.00
39 T	Methylcyclohexane	0.585	0.616	-5.3	89	0.00
40 TM	Benzene	1.417	1.463	-3.2	89	0.00
41 T	Methacrylonitrile	0.258	0.250	3.1	83	-0.03
42 TM	1,2-Dichloroethane	0.512	0.521	-1.8	89	-0.01
43 T	Isopropyl Acetate	0.810	0.781	3.6	81	0.00
44 TM	Trichloroethene	0.395	0.406	-2.8	92	0.00
45 C	1,2-Dichloropropane	0.368	0.375	-1.9#	89	0.00
46 T	Dibromomethane	0.265	0.281	-6.0	90	0.00
47 T	Bromodichloromethane	0.488	0.519	-6.4	88	0.00
48 T	Methyl methacrylate	0.387	0.378	2.3	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092923\
 Data File : VX037989.D
 Acq On : 29 Sep 2023 20:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 29 23:11:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.009	10.0	82	-0.01
50 S	Toluene-d8	1.243	1.286	-3.5	86	0.00
51 T	4-Methyl-2-Pentanone	0.514	0.497	3.3	82	0.00
52 CM	Toluene	0.913	0.981	-7.4#	90	0.00
53 T	t-1,3-Dichloropropene	0.510	0.556	-9.0	88	-0.01
54 T	cis-1,3-Dichloropropene	0.564	0.602	-6.7	88	0.00
55 T	1,1,2-Trichloroethane	0.385	0.417	-8.3	92	0.00
56 T	Ethyl methacrylate	0.575	0.599	-4.2	86	0.00
57 T	1,3-Dichloropropane	0.623	0.652	-4.7	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.268	0.214	20.1	69	0.00
59 T	2-Hexanone	0.400	0.382	4.5	81	0.00
60 T	Dibromochloromethane	0.384	0.431	-12.2	91	0.00
61 T	1,2-Dibromoethane	0.408	0.449	-10.0	92	0.00
62 S	4-Bromofluorobenzene	0.452	0.513	-13.5	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.341	0.367	-7.6	97	0.00
65 PM	Chlorobenzene	1.079	1.129	-4.6	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.382	0.417	-9.2	92	0.00
67 C	Ethyl Benzene	1.870	1.995	-6.7#	93	0.00
68 T	m/p-Xylenes	0.723	0.781	-8.0	93	0.00
69 T	o-Xylene	0.725	0.763	-5.2	91	0.00
70 T	Styrene	1.163	1.298	-11.6	94	0.00
71 P	Bromoform	0.302	0.330	-9.3	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.668	3.477	5.2	93	0.00
74 T	N-amyl acetate	1.486	1.333	10.3	87	0.00
75 P	1,1,2,2-Tetrachloroethane	1.290	1.180	8.5	94	0.00
76 T	1,2,3-Trichloropropane	1.156	1.031	10.8	89	0.00
77 T	Bromobenzene	0.908	0.876	3.5	97	0.00
78 T	n-propylbenzene	4.143	4.130	0.3	96	0.00
79 T	2-Chlorotoluene	2.554	2.402	6.0	94	0.00
80 T	1,3,5-Trimethylbenzene	3.098	3.008	2.9	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.361	0.312	13.6	82	0.00
82 T	4-Chlorotoluene	2.876	2.833	1.5	97	0.00
83 T	tert-Butylbenzene	3.131	2.990	4.5	93	0.00
84 T	1,2,4-Trimethylbenzene	3.082	3.013	2.2	95	0.00
85 T	sec-Butylbenzene	3.798	3.829	-0.8	97	0.00
86 T	p-Isopropyltoluene	3.165	3.244	-2.5	99	0.00
87 T	1,3-Dichlorobenzene	1.692	1.731	-2.3	104	0.00
88 T	1,4-Dichlorobenzene	1.732	1.773	-2.4	107	0.00
89 T	n-Butylbenzene	2.772	3.016	-8.8	104	0.00
90 T	Hexachloroethane	0.536	0.534	0.4	95	0.00
91 T	1,2-Dichlorobenzene	1.728	1.722	0.3	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.241	6.2	90	0.00
93 T	1,2,4-Trichlorobenzene	1.018	1.167	-14.6	114	0.00
94 T	Hexachlorobutadiene	0.475	0.509	-7.2	106	0.00
95 T	Naphthalene	3.442	3.597	-4.5	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092923\
Data File : VX037989.D
Acq On : 29 Sep 2023 20:24
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 29 23:11:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.047	1.144	-9.3	109	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6