

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
 Data File : VX047874.D
 Acq On : 29 Sep 2025 11:48
 Operator : JC/MD
 Sample : Q3156-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB03-20250912

Quant Time: Sep 30 01:17:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	122948	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	260114	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	260803	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	133062	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	111153	56.797	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 113.600%	
35) Dibromofluoromethane	5.379	113	93369	53.523	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.040%	
50) Toluene-d8	8.635	98	312358	51.747	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.500%	
62) 4-Bromofluorobenzene	11.067	95	132327	57.763	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 115.520%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
Data File : VX047874.D
Acq On : 29 Sep 2025 11:48
Operator : JC/MD
Sample : Q3156-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB03-20250912

Quant Time: Sep 30 01:17:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

