

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
 Data File : VX047875.D
 Acq On : 29 Sep 2025 12:09
 Operator : JC/MD
 Sample : Q3156-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB03-20250918

Quant Time: Sep 30 01:18:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	191712	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	390271	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	393970	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	200279	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	149465	48.980	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.960%
35) Dibromofluoromethane	5.373	113	126244	48.233	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.460%
50) Toluene-d8	8.634	98	446748	49.328	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.660%
62) 4-Bromofluorobenzene	11.067	95	186040	54.126	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.260%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	15701	11.833	ug/l	98
25) 2-Butanone	4.556	43	9300	5.619	ug/l	96
29) Tetrahydrofuran	5.001	42	9392	9.433	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
Data File : VX047875.D
Acq On : 29 Sep 2025 12:09
Operator : JC/MD
Sample : Q3156-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB03-20250918

Quant Time: Sep 30 01:18:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

