

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
 Data File : VX047876.D
 Acq On : 29 Sep 2025 12:30
 Operator : JC/MD
 Sample : Q3156-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB02-20250918

Quant Time: Sep 30 01:18:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	147800	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	298872	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	312582	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	158720	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	124056	52.731	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.460%	
35) Dibromofluoromethane	5.379	113	100561	50.170	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.340%	
50) Toluene-d8	8.635	98	339834	48.998	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.000%	
62) 4-Bromofluorobenzene	11.067	95	145925	55.439	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 110.880%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	14651	14.323	ug/l	95
29) Tetrahydrofuran	5.013	42	3950	5.146	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
Data File : VX047876.D
Acq On : 29 Sep 2025 12:30
Operator : JC/MD
Sample : Q3156-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB02-20250918

Quant Time: Sep 30 01:18:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

