

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
 Data File : VX047878.D
 Acq On : 29 Sep 2025 13:13
 Operator : JC/MD
 Sample : Q3156-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT191D1-20250918

Quant Time: Sep 30 01:19:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	174530	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	361493	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	366051	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	179524	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	149648	53.867	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.740%	
35) Dibromofluoromethane	5.379	113	127534	52.605	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.220%	
50) Toluene-d8	8.635	98	441646	52.647	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.300%	
62) 4-Bromofluorobenzene	11.067	95	183690	57.697	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 115.400%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2412	1.997	ug/l #	63
20) Methylene Chloride	2.794	84	1050	0.411	ug/l #	74
44) Trichloroethene	7.117	130	2116	0.813	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
Data File : VX047878.D
Acq On : 29 Sep 2025 13:13
Operator : JC/MD
Sample : Q3156-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT191D1-20250918

Quant Time: Sep 30 01:19:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

