

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
 Data File : VX047879.D
 Acq On : 29 Sep 2025 13:34
 Operator : JC/MD
 Sample : Q3156-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT180D1-20250918

Quant Time: Sep 30 01:20:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	176708	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	368400	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	368560	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	180399	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	152938	54.373	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.740%	
35) Dibromofluoromethane	5.373	113	130216	52.704	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.400%	
50) Toluene-d8	8.635	98	448703	52.485	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.980%	
62) 4-Bromofluorobenzene	11.067	95	189490	58.403	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 116.800%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2074	1.696	ug/l #	82
44) Trichloroethene	7.110	130	11909	4.492	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
Data File : VX047879.D
Acq On : 29 Sep 2025 13:34
Operator : JC/MD
Sample : Q3156-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT180D1-20250918

Quant Time: Sep 30 01:20:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

