

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
 Data File : VX047885.D
 Acq On : 29 Sep 2025 15:41
 Operator : JC/MD
 Sample : Q3213-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-TB-20250923

Quant Time: Sep 30 01:22:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	141653	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.750	114	299169	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.036	117	307943	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	161425	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	124215	55.090	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.180%
35) Dibromofluoromethane	5.379	113	101921	50.798	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.600%
50) Toluene-d8	8.634	98	340143	48.994	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.980%
62) 4-Bromofluorobenzene	11.067	95	148173	56.237	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
Data File : VX047885.D
Acq On : 29 Sep 2025 15:41
Operator : JC/MD
Sample : Q3213-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-TB-20250923

Quant Time: Sep 30 01:22:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

