

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
 Data File : VX047886.D
 Acq On : 29 Sep 2025 16:02
 Operator : JC/MD
 Sample : Q3213-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-GW-340-342

Quant Time: Sep 30 01:22:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	121262	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	250493	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	252342	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	126414	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	96623	50.059	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.120%	
35) Dibromofluoromethane	5.379	113	80030	47.639	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.280%	
50) Toluene-d8	8.635	98	278316	47.879	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 95.760%	
62) 4-Bromofluorobenzene	11.067	95	113538	51.465	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.940%	
Target Compounds						
						Qvalue
3) Chloromethane	1.307	50	815	0.494	ug/l	93
16) Acetone	2.380	43	5987	7.134	ug/l	98
17) Carbon Disulfide	2.520	76	1904	0.483	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
Data File : VX047886.D
Acq On : 29 Sep 2025 16:02
Operator : JC/MD
Sample : Q3213-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-GW-340-342

Quant Time: Sep 30 01:22:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

