

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX092925\  
 Data File : VX047887.D  
 Acq On : 29 Sep 2025 16:23  
 Operator : JC/MD  
 Sample : Q3213-04  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BP-VPB-184-GW-360-362

Quant Time: Sep 30 01:23:24 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091625W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 17 06:39:58 2025  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| -----                       |        |       |          |          |            |          |
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.544  | 168   | 147966   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.751  | 114   | 308643   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.037 | 117   | 320742   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.006 | 152   | 164731   | 50.000   | ug/l       | 0.00     |
|                             |        |       |          |          |            |          |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.946  | 65    | 130174   | 55.270   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | = 110.540% |          |
| 35) Dibromofluoromethane    | 5.379  | 113   | 105173   | 50.810   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | = 101.620% |          |
| 50) Toluene-d8              | 8.635  | 98    | 353678   | 49.380   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | = 98.760%  |          |
| 62) 4-Bromofluorobenzene    | 11.067 | 95    | 156453   | 57.557   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | = 115.120% |          |
|                             |        |       |          |          |            |          |
| Target Compounds            |        |       |          |          |            | Qvalue   |
| 16) Acetone                 | 2.380  | 43    | 14228    | 13.894   | ug/l       | 96       |
| 25) 2-Butanone              | 4.574  | 43    | 3897     | 3.051    | ug/l #     | 81       |
| -----                       |        |       |          |          |            |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX092925\  
Data File : VX047887.D  
Acq On : 29 Sep 2025 16:23  
Operator : JC/MD  
Sample : Q3213-04  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BP-VPB-184-GW-360-362

Quant Time: Sep 30 01:23:24 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X091625W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 17 06:39:58 2025  
Response via : Initial Calibration

