

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
 Data File : VX047888.D
 Acq On : 29 Sep 2025 16:44
 Operator : JC/MD
 Sample : Q3213-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-GW-380-382

Quant Time: Sep 30 01:23:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	133701	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	285495	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	296119	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	153022	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	121464	57.074	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 114.140%	
35) Dibromofluoromethane	5.379	113	96700	50.504	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.000%	
50) Toluene-d8	8.635	98	321663	48.551	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.100%	
62) 4-Bromofluorobenzene	11.061	95	141912	56.440	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 112.880%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.301	50	1081	0.594	ug/l	96
16) Acetone	2.374	43	5941	6.420	ug/l	99
17) Carbon Disulfide	2.520	76	1044	0.240	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
Data File : VX047888.D
Acq On : 29 Sep 2025 16:44
Operator : JC/MD
Sample : Q3213-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-GW-380-382

Quant Time: Sep 30 01:23:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

