

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
 Data File : VX047891.D
 Acq On : 29 Sep 2025 17:46
 Operator : JC/MD
 Sample : Q3213-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-GW-440-442

Quant Time: Sep 30 01:25:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	126934	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	264916	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	273054	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	137102	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	105715	52.322	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.640%
35) Dibromofluoromethane	5.373	113	88669	49.907	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.820%
50) Toluene-d8	8.635	98	293810	47.792	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.580%
62) 4-Bromofluorobenzene	11.061	95	125076	53.608	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.220%
Target Compounds						
						Qvalue
3) Chloromethane	1.300	50	789	0.457	ug/l	98
16) Acetone	2.380	43	4779	5.440	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092925\
Data File : VX047891.D
Acq On : 29 Sep 2025 17:46
Operator : JC/MD
Sample : Q3213-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-GW-440-442

Quant Time: Sep 30 01:25:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

