

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX093019\
 Data File : VX012702.D
 Acq On : 30 Sep 2019 13:59
 Operator : JC/SP
 Sample : K5081-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-PZ183I-20190925

Quant Time: Oct 01 04:20:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	168136	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	291426	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	236509	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	82350	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	119470	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
35) Dibromofluoromethane	5.49	113	87540	50.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.88%	
50) Toluene-d8	8.71	98	337698	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
62) 4-Bromofluorobenzene	11.13	95	108008	41.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.20%	
Target Compounds						
16) Acetone	2.43	43	11875	9.157	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX093019\
Data File : VX012702.D
Acq On : 30 Sep 2019 13:59
Operator : JC/SP
Sample : K5081-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-PZ183I-20190925

Quant Time: Oct 01 04:20:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

