

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX093019\
 Data File : VX012703.D
 Acq On : 30 Sep 2019 14:22
 Operator : JC/SP
 Sample : K5081-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-PZ182I-20190925

Quant Time: Oct 01 04:24:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	169498	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	295858	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	249787	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	96931	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119967	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
35) Dibromofluoromethane	5.49	113	90426	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
50) Toluene-d8	8.71	98	347426	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
62) 4-Bromofluorobenzene	11.13	95	117757	44.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.28%	
Target Compounds						
16) Acetone	2.43	43	11610	8.880	ug/l	96
24) 1,1-Dichloroethane	3.69	63	7208	1.898	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX093019\
Data File : VX012703.D
Acq On : 30 Sep 2019 14:22
Operator : JC/SP
Sample : K5081-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-PZ182I-20190925

Quant Time: Oct 01 04:24:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

