

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX093019\
 Data File : VX012704.D
 Acq On : 30 Sep 2019 14:46
 Operator : JC/SP
 Sample : K5081-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-DUP03-20190925

Quant Time: Oct 01 04:26:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	163411	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	281450	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	237434	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	87150	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	115208	48.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.84%	
35) Dibromofluoromethane	5.48	113	85727	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
50) Toluene-d8	8.71	98	332201	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
62) 4-Bromofluorobenzene	11.14	95	111840	44.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.12%	
Target Compounds						
12) 1,1-Dichloroethene	2.36	96	1862	1.060	ug/l	79
16) Acetone	2.43	43	11002	8.729	ug/l	93
24) 1,1-Dichloroethane	3.69	63	39283	10.729	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX093019\
Data File : VX012704.D
Acq On : 30 Sep 2019 14:46
Operator : JC/SP
Sample : K5081-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-DUP03-20190925

Quant Time: Oct 01 04:26:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

