

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX093019\
 Data File : VX012709.D
 Acq On : 30 Sep 2019 16:43
 Operator : JC/SP
 Sample : K5081-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-PZ12311-20190926

Quant Time: Oct 01 04:36:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	165460	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	286417	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	242402	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	92934	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	115694	48.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.04%	
35) Dibromofluoromethane	5.49	113	85155	49.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.84%	
50) Toluene-d8	8.71	98	336166	51.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.36%	
62) 4-Bromofluorobenzene	11.14	95	115312	44.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.28%	
Target Compounds						
16) Acetone	2.44	43	4068	3.188	ug/l	92
24) 1,1-Dichloroethane	3.69	63	4753	1.282	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX093019\
Data File : VX012709.D
Acq On : 30 Sep 2019 16:43
Operator : JC/SP
Sample : K5081-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-PZ123I1-20190926

Quant Time: Oct 01 04:36:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

