

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX093019\
 Data File : VX012718.D
 Acq On : 30 Sep 2019 20:13
 Operator : JC/SP
 Sample : K5081-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FT-DUP04-20190925

Quant Time: Oct 01 10:37:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	169556	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	288586	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	237441	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	81738	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	116179	47.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.10%	
35) Dibromofluoromethane	5.49	113	86289	50.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.42%	
50) Toluene-d8	8.71	98	337355	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
62) 4-Bromofluorobenzene	11.14	95	109939	42.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.48%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	1448	0.778	ug/l	94
16) Acetone	2.43	43	10202	7.801	ug/l #	83
24) 1,1-Dichloroethane	3.69	63	5479	1.442	ug/l #	94
32) 1,1,1-Trichloroethane	5.49	97	11379	3.303	ug/l	99
44) Trichloroethene	7.22	130	1766	0.844	ug/l	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX093019\
Data File : VX012718.D
Acq On : 30 Sep 2019 20:13
Operator : JC/SP
Sample : K5081-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FT-DUP04-20190925

Quant Time: Oct 01 10:37:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

