

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX093019\
 Data File : VX012719.D
 Acq On : 30 Sep 2019 20:36
 Operator : JC/SP
 Sample : K5081-22
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-MW126S-20190926

Quant Time: Oct 01 04:54:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	166191	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	287781	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	240093	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	89516	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	117223	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
35) Dibromofluoromethane	5.49	113	87025	50.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.56%	
50) Toluene-d8	8.71	98	336186	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
62) 4-Bromofluorobenzene	11.14	95	112376	43.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.60%	
Target Compounds						
16) Acetone	2.44	43	7097	5.536	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX093019\
Data File : VX012719.D
Acq On : 30 Sep 2019 20:36
Operator : JC/SP
Sample : K5081-22
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-MW126S-20190926

Quant Time: Oct 01 04:54:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

