

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VOX093020\
 Data File : VX018702.D
 Acq On : 30 Sep 2020 17:11
 Operator : JC/SP
 Sample : L4135-07DL2 200X
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BG481DL2

Quant Time: Oct 01 04:03:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR092820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 01 01:06:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	229856	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	222222	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	100317	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	41310	3.43	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.60%
7) Chloroethane-d5	1.70	69	39192	3.86	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.20%
11) 1,1-Dichloroethene-d2	2.34	63	82941	3.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.40%
20) 2-Butanone-d5	4.56	46	165682	54.02	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.04%
24) Chloroform-d	5.14	84	128747	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.40%
26) 1,2-Dichloroethane-d4	6.03	65	72277	4.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.00%
32) Benzene-d6	6.06	84	210814	4.11	ug/L	0.01
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.20%
36) 1,2-Dichloropropane-d6	7.37	67	69455	4.44	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.80%
41) Toluene-d8	8.70	98	185021	3.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.60%
45) trans-1,3-Dichloropropene-	8.99	79	29667	4.27	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.40%
48) 2-Hexanone-d5	9.43	63	121054	48.63	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.26%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	55757	4.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.80%
65) 1,2-Dichlorobenzene-d4	12.36	152	75912	4.63	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.60%

Target Compounds					Ovalue
16) Methylene chloride	2.84	84	5456	0.356 ug/L	85
53) Chlorobenzene	10.12	112	261245	8.545 ug/L	98
58) Isopropylbenzene	11.00	105	82058	1.680 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX093020\
Data File : VX018702.D
Acq On : 30 Sep 2020 17:11
Operator : JC/SP
Sample : L4135-07DL2 200X
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BG481DL2

Quant Time: Oct 01 04:03:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR092820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 01 01:06:26 2020
Response via : Initial Calibration

