



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093022\
 Data File : VX031805.D
 Acq On : 30 Sep 2022 08:14
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X093022W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Mon Oct 03 00:56:55 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.1	4251	PASS
75	95	30	60	49.4	12307	PASS
95	95	100	100	100.0	24915	PASS
96	95	5	9	6.8	1699	PASS
173	174	0.00	2	1.3	254	PASS
174	95	50	100	79.8	19886	PASS
175	174	5	9	6.4	1265	PASS
176	174	95	101	96.8	19244	PASS
177	176	5	9	6.1	1178	PASS