

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093023\
 Data File : VX037996.D
 Acq On : 30 Sep 2023 12:12
 Operator : JC/MD
 Sample : 04651-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP1-3

Quant Time: Oct 02 02:08:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	212480	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	311652	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	366748	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	158891	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	133920	47.971	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.940%
35) Dibromofluoromethane	5.391	113	115140	55.216	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.440%
50) Toluene-d8	8.647	98	451321	58.249	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	116.500%#
62) 4-Bromofluorobenzene	11.079	95	145988	51.851	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.700%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	27250	20.291	ug/l	91
20) Methylene Chloride	2.788	84	14802	5.939	ug/l	87
43) Isopropyl Acetate	6.342	43	126219	24.996	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093023\
Data File : VX037996.D
Acq On : 30 Sep 2023 12:12
Operator : JC/MD
Sample : 04651-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP1-3

Quant Time: Oct 02 02:08:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

