

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093023\
 Data File : VX038002.D
 Acq On : 30 Sep 2023 14:31
 Operator : JC/MD
 Sample : 04652-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IS6-1

Quant Time: Oct 02 02:11:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	146507	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	259011	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	269446	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139949	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	108559	56.398	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.800%
35) Dibromofluoromethane	5.391	113	84598	48.814	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.620%
50) Toluene-d8	8.647	98	325648	50.571	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.140%
62) 4-Bromofluorobenzene	11.079	95	129859	55.496	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.000%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	13694	14.789	ug/l	95
20) Methylene Chloride	2.794	84	7685	4.466	ug/l	93
43) Isopropyl Acetate	6.342	43	74268	17.697	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093023\
Data File : VX038002.D
Acq On : 30 Sep 2023 14:31
Operator : JC/MD
Sample : 04652-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IS6-1

Quant Time: Oct 02 02:11:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

