

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093023\
 Data File : VX038003.D
 Acq On : 30 Sep 2023 14:54
 Operator : JC/MD
 Sample : 04666-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-201

Quant Time: Oct 02 02:11:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	147184	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	263937	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	272952	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139243	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	109399	56.573	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 113.140%	
35) Dibromofluoromethane	5.391	113	85670	48.510	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.020%	
50) Toluene-d8	8.647	98	330449	50.359	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.720%	
62) 4-Bromofluorobenzene	11.079	95	130037	54.535	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.080%	
Target Compounds						Qvalue
16) Acetone	2.380	43	12933	13.903	ug/l	91
20) Methylene Chloride	2.794	84	7323	4.235	ug/l	91
43) Isopropyl Acetate	6.342	43	83834	19.603	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093023\
Data File : VX038003.D
Acq On : 30 Sep 2023 14:54
Operator : JC/MD
Sample : 04666-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-201

Quant Time: Oct 02 02:11:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

