

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093023\
 Data File : VX038005.D
 Acq On : 30 Sep 2023 15:40
 Operator : JC/MD
 Sample : 04647-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S-2

Quant Time: Oct 02 02:12:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	152423	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	272577	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	282261	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	147987	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	113477	56.665	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.320%
35) Dibromofluoromethane	5.385	113	89333	48.981	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.960%
50) Toluene-d8	8.647	98	344775	50.877	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.760%
62) 4-Bromofluorobenzene	11.079	95	137778	55.950	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.900%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	25936	26.922	ug/l	94
20) Methylene Chloride	2.788	84	8667	4.843	ug/l #	88
40) Benzene	6.044	78	42048	5.442	ug/l	98
43) Isopropyl Acetate	6.342	43	91672	20.757	ug/l	98
52) Toluene	8.720	92	35481	7.126	ug/l	94
68) m/p-Xylenes	10.299	106	13290	3.258	ug/l	98
69) o-Xylene	10.640	106	8724	2.133	ug/l	100
95) Naphthalene	13.774	128	95994	9.422	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093023\
Data File : VX038005.D
Acq On : 30 Sep 2023 15:40
Operator : JC/MD
Sample : 04647-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S-2

Quant Time: Oct 02 02:12:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

