

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093023\
 Data File : VX038015.D
 Acq On : 30 Sep 2023 19:31
 Operator : JC/MD
 Sample : 04651-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP1-3

Quant Time: Oct 02 02:17:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	139884	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	253499	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	268254	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136150	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	104786	57.015	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.040%
35) Dibromofluoromethane	5.385	113	79222	46.706	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.420%
50) Toluene-d8	8.647	98	332247	52.718	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.440%
62) 4-Bromofluorobenzene	11.079	95	128479	56.101	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.200%
Target Compounds						Qvalue
16) Acetone	2.380	43	20451	23.131	ug/l	100
20) Methylene Chloride	2.794	84	9165	5.585	ug/l	91
43) Isopropyl Acetate	6.342	43	93391	22.737	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093023\
Data File : VX038015.D
Acq On : 30 Sep 2023 19:31
Operator : JC/MD
Sample : 04651-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP1-3

Quant Time: Oct 02 02:17:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

