

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093023\
 Data File : VX038016.D
 Acq On : 30 Sep 2023 19:54
 Operator : JC/MD
 Sample : 04652-05RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IS2-3RE

Quant Time: Oct 02 02:18:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	249267	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	493673	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	467160	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	275154	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	174942	53.418	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.840%
35) Dibromofluoromethane	5.385	113	142057	43.006	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.020%
50) Toluene-d8	8.647	98	519565	42.333	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	84.660%#
62) 4-Bromofluorobenzene	11.079	95	222134	49.807	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.620%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	18661	11.845	ug/l	90
20) Methylene Chloride	2.794	84	12723	4.345	ug/l	90
43) Isopropyl Acetate	6.342	43	116972	14.624	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093023\
Data File : VX038016.D
Acq On : 30 Sep 2023 19:54
Operator : JC/MD
Sample : 04652-05RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IS2-3RE

Quant Time: Oct 02 02:18:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

