

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093024\
 Data File : VX043197.D
 Acq On : 30 Sep 2024 14:42
 Operator : JC/MD
 Sample : PB163747TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB163747TB

Quant Time: Oct 01 01:52:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	90147	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	176558	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	150053	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66254	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	78251	52.258	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.520%			
35) Dibromofluoromethane	5.385	113	59071	47.428	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 94.860%			
50) Toluene-d8	8.647	98	208218	49.413	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.820%			
62) 4-Bromofluorobenzene	11.079	95	75598	44.549	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 89.100%			
Target Compounds						
16) Acetone	2.392	43	32337	46.178	ug/l	98
20) Methylene Chloride	2.782	84	3390	3.043	ug/l	91
25) 2-Butanone	4.587	43	4603	5.099	ug/l	99
43) Isopropyl Acetate	6.342	43	88301	27.501	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093024\
Data File : VX043197.D
Acq On : 30 Sep 2024 14:42
Operator : JC/MD
Sample : PB163747TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB163747TB

Quant Time: Oct 01 01:52:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration

