

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093024\
 Data File : VX043198.D
 Acq On : 30 Sep 2024 15:05
 Operator : JC/MD
 Sample : P4214-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CONCRETE-WC-1

Quant Time: Oct 01 09:31:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	95372	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	185065	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63157	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83089	52.449	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.900%			
35) Dibromofluoromethane	5.385	113	33819	25.905	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 51.800%#			
50) Toluene-d8	8.647	98	217250	49.187	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.380%			
62) 4-Bromofluorobenzene	11.079	95	75205	42.280	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 84.560%			
Target Compounds						
16) Acetone	2.386	43	82393	111.213	ug/l	99
18) Methyl Acetate	2.703	43	6229	3.124	ug/l	99
20) Methylene Chloride	2.788	84	2641	2.241	ug/l #	92
25) 2-Butanone	4.580	43	4788	5.014	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093024\
Data File : VX043198.D
Acq On : 30 Sep 2024 15:05
Operator : JC/MD
Sample : P4214-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CONCRETE-WC-1

Quant Time: Oct 01 09:31:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration

