

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093024\
 Data File : VX043199.D
 Acq On : 30 Sep 2024 15:28
 Operator : JC/MD
 Sample : P4222-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-5

Quant Time: Oct 01 01:53:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	96561	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	186462	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	157859	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68423	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83032	51.768	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.540%
35) Dibromofluoromethane	5.385	113	62823	47.761	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.520%
50) Toluene-d8	8.647	98	221210	49.708	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.420%
62) 4-Bromofluorobenzene	11.079	95	78807	43.974	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	87.940%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	25445	33.922	ug/l	97
18) Methyl Acetate	2.703	43	12106	5.996	ug/l	98
20) Methylene Chloride	2.788	84	2488	2.085	ug/l #	92
43) Isopropyl Acetate	6.342	43	87604	25.835	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093024\
Data File : VX043199.D
Acq On : 30 Sep 2024 15:28
Operator : JC/MD
Sample : P4222-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-5

Quant Time: Oct 01 01:53:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration

