

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100118\
 Data File : VX004879.D
 Acq On : 01 Oct 2018 19:03
 Operator : JC/MD
 Sample : J5135-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW508-3337

Manual Integrations
 APPROVED

MMDadoda
 10/4/2018 7:07:52 PM

Quant Time: Oct 11 13:45:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	202987	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	336231	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	333672	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	184876	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	164577	56.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.52%	
35) Dibromofluoromethane	5.50	113	139577	48.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.96%	
50) Toluene-d8	8.72	98	496441	52.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.80%	
62) 4-Bromofluorobenzene	11.14	95	174935	51.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.52%	
Target Compounds						
					Qvalue	
6) Chloroethane	1.73	64	106139	62.419	ug/l	97
12) 1,1-Dichloroethene	2.37	96	65744	29.891	ug/l	92
16) Acetone	2.45	43	3192	1.886	ug/l	93
24) 1,1-Dichloroethane	3.70	63	2240375	464.765	ug/l	98
32) 1,1,1-Trichloroethane	5.49	97	55633	14.953	ug/l	99
40) Benzene	6.14	78	9801	0.805	ug/l	100
44) Trichloroethene	7.21	130	1268	0.424	ug/l	84
85) sec-Butylbenzene	11.94	105	5676	0.486	ug/l #	77
88) 1,4-Dichlorobenzene	12.10	146	8735m	1.339	ug/l	
93) 1,2,4-Trichlorobenzene	13.65	180	16166	3.489	ug/l	96
96) 1,2,3-Trichlorobenzene	14.02	180	5335	1.115	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100118\
Data File : VX004879.D
Acq On : 01 Oct 2018 19:03
Operator : JC/MD
Sample : J5135-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW508-3337

Manual Integrations
APPROVED

MMDadoda
10/4/2018 7:07:52 PM

Quant Time: Oct 11 13:45:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

