

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100118\
 Data File : VX004890.D
 Acq On : 02 Oct 2018 00:51
 Operator : JC/MD
 Sample : J5168-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 QNWP8-TRIP-BLANK-2

Quant Time: Oct 03 13:18:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	204337	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	339390	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	336194	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	183000	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	164281	56.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.56%	
35) Dibromofluoromethane	5.50	113	138714	48.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.46%	
50) Toluene-d8	8.72	98	494576	51.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.44%	
62) 4-Bromofluorobenzene	11.14	95	175444	50.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.86%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100118\
Data File : VX004890.D
Acq On : 02 Oct 2018 00:51
Operator : JC/MD
Sample : J5168-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
QNWP8-TRIP-BLANK-2

Quant Time: Oct 03 13:18:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

