

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100118\
 Data File : VX004895.D
 Acq On : 02 Oct 2018 03:04
 Operator : JC/MD
 Sample : J5041-03DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S13-0251DL

Quant Time: Oct 03 13:23:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	220076	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	351399	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	345648	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	203299	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	167119	53.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.32%	
35) Dibromofluoromethane	5.50	113	139228	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
50) Toluene-d8	8.72	98	520253	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
62) 4-Bromofluorobenzene	11.14	95	191882	53.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.60%	
Target Compounds						
					Qvalue	
31) Cyclohexane	5.57	56	67205	16.801	ug/l	95
39) Methylcyclohexane	7.46	83	76522	18.443	ug/l	93
67) Ethyl Benzene	10.26	91	178172	12.662	ug/l	98
73) Isopropylbenzene	11.02	105	36183	2.824	ug/l	99
78) n-propylbenzene	11.36	91	48741	3.315	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	41919	3.385	ug/l	96
84) 1,2,4-Trimethylbenzene	11.81	105	275150	22.609	ug/l	98
95) Naphthalene	13.83	128	124784	7.992	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100118\
Data File : VX004895.D
Acq On : 02 Oct 2018 03:04
Operator : JC/MD
Sample : J5041-03DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleID :
S13-0251DL

Quant Time: Oct 03 13:23:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

