

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100119\
 Data File : VX012727.D
 Acq On : 01 Oct 2019 12:54
 Operator : JC/SP
 Sample : K5081-10RE
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FT-DUP04-20190925RE

Quant Time: Oct 02 11:32:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	164164	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	283734	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	237209	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	80310	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	115441	48.29	ug/l	0.00
Spiked Amount			Recovery	=	96.58%	
35) Dibromofluoromethane	5.49	113	86261	51.05	ug/l	0.00
Spiked Amount			Recovery	=	102.10%	
50) Toluene-d8	8.71	98	334108	51.34	ug/l	0.00
Spiked Amount			Recovery	=	102.68%	
62) 4-Bromofluorobenzene	11.14	95	105089	41.07	ug/l	0.00
Spiked Amount			Recovery	=	82.14%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.37	101	1120	0.621	ug/l #	77
16) Acetone	2.44	43	11161	8.814	ug/l #	89
24) 1,1-Dichloroethane	3.69	63	5293	1.439	ug/l	97
32) 1,1,1-Trichloroethane	5.49	97	11437	3.429	ug/l	99
44) Trichloroethene	7.20	130	1607	0.781	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100119\
Data File : VX012727.D
Acq On : 01 Oct 2019 12:54
Operator : JC/SP
Sample : K5081-10RE
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FT-DUP04-20190925RE

Quant Time: Oct 02 11:32:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

