

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100119\
 Data File : VX012729.D
 Acq On : 01 Oct 2019 13:41
 Operator : JC/SP
 Sample : K5081-14RE
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-PZ123I-20190926RE

Quant Time: Oct 02 03:43:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	159274	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	269447	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	217513	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	71556	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	108011	46.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.14%	
35) Dibromofluoromethane	5.49	113	82347	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
50) Toluene-d8	8.71	98	312418	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
62) 4-Bromofluorobenzene	11.14	95	97440	40.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.20%	
Target Compounds						
12) 1,1-Dichloroethene	2.36	96	10498	6.130	ug/l #	95
16) Acetone	2.44	43	10974	8.933	ug/l	99
24) 1,1-Dichloroethane	3.69	63	196320	55.014	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100119\
Data File : VX012729.D
Acq On : 01 Oct 2019 13:41
Operator : JC/SP
Sample : K5081-14RE
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-PZ123I-20190926RE

Quant Time: Oct 02 03:43:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

