

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100119\
 Data File : VX012732.D
 Acq On : 01 Oct 2019 14:51
 Operator : JC/SP
 Sample : K5082-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-MW126D-20190926

Quant Time: Oct 02 05:45:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	163448	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	282616	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	242824	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	89564	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	115399	48.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.96%	
35) Dibromofluoromethane	5.49	113	85401	50.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.48%	
50) Toluene-d8	8.71	98	335643	51.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.56%	
62) 4-Bromofluorobenzene	11.14	95	110711	43.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.88%	
Target Compounds						
16) Acetone	2.43	43	7713	6.118	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100119\
Data File : VX012732.D
Acq On : 01 Oct 2019 14:51
Operator : JC/SP
Sample : K5082-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-MW126D-20190926

Quant Time: Oct 02 05:45:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

