

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100119\
 Data File : VX012740.D
 Acq On : 01 Oct 2019 17:58
 Operator : JC/SP
 Sample : K5082-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-SW204-20190926

Quant Time: Oct 02 04:32:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	159898	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	281819	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	233639	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	81422	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	113400	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
35) Dibromofluoromethane	5.49	113	83072	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
50) Toluene-d8	8.71	98	327305	50.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.28%	
62) 4-Bromofluorobenzene	11.14	95	103819	40.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.70%	
Target Compounds						
16) Acetone	2.43	43	8376	6.791	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100119\
Data File : VX012740.D
Acq On : 01 Oct 2019 17:58
Operator : JC/SP
Sample : K5082-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-SW204-20190926

Quant Time: Oct 02 04:32:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

