

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100119\
 Data File : VX012743.D
 Acq On : 01 Oct 2019 19:08
 Operator : JC/SP
 Sample : K5110-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Oct 02 08:50:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	163879	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	287000	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	240857	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	93696	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	120932	50.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.36%	
35) Dibromofluoromethane	5.49	113	87903	51.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.86%	
50) Toluene-d8	8.71	98	343172	52.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.28%	
62) 4-Bromofluorobenzene	11.14	95	116517	45.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.04%	
Target Compounds						
12) 1,1-Dichloroethene	2.36	96	1036	0.588	ug/l	97
16) Acetone	2.43	43	3201	2.532	ug/l	91
27) cis-1,2-Dichloroethene	4.59	96	1409	0.622	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100119\
Data File : VX012743.D
Acq On : 01 Oct 2019 19:08
Operator : JC/SP
Sample : K5110-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Oct 02 08:50:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

