

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100119\
 Data File : VX012745.D
 Acq On : 01 Oct 2019 19:55
 Operator : JC/SP
 Sample : K5110-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3

Quant Time: Oct 02 05:56:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	161932	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	285356	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	236229	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	85102	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	117471	49.82	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.64%	
35) Dibromofluoromethane	5.49	113	88561	52.11	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	104.22%	
50) Toluene-d8	8.71	98	342734	52.37	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	104.74%	
62) 4-Bromofluorobenzene	11.14	95	109252	42.46	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	84.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	7560	3.636	ug/l	93
12) 1,1-Dichloroethene	2.36	96	30457	17.493	ug/l	99
16) Acetone	2.43	43	10530	8.431	ug/l	99
24) 1,1-Dichloroethane	3.69	63	19776	5.451	ug/l	98
27) cis-1,2-Dichloroethene	4.59	96	7753	3.464	ug/l	91
40) Benzene	6.14	78	3118	0.385	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100119\
Data File : VX012745.D
Acq On : 01 Oct 2019 19:55
Operator : JC/SP
Sample : K5110-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3

Quant Time: Oct 02 05:56:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

