

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100120\
 Data File : VX018729.D
 Acq On : 01 Oct 2020 17:46
 Operator : JC/SP
 Sample : VX1001WBL01
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK76

Quant Time: Oct 02 09:07:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR092820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 02 08:54:14 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	215345	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	208192	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	94817	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	63742	5.66	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	113.20%
7) Chloroethane-d5	1.70	69	49193	5.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.60%
11) 1,1-Dichloroethene-d2	2.34	63	109233	4.39	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.80%
20) 2-Butanone-d5	4.56	46	161644	56.25	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.50%
24) Chloroform-d	5.14	84	137324	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.80%
26) 1,2-Dichloroethane-d4	6.03	65	85205	5.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.60%
32) Benzene-d6	6.05	84	242779	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.00%
36) 1,2-Dichloropropane-d6	7.37	67	69769	4.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.20%
41) Toluene-d8	8.70	98	230515	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.40%
45) trans-1,3-Dichloropropene-	8.99	79	34447	5.29	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	105.80%
48) 2-Hexanone-d5	9.43	63	126198	54.11	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.22%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	54233	4.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.40%
65) 1,2-Dichlorobenzene-d4	12.36	152	89553	5.78	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.60%

Target Compounds					Ovalue
16) Methylene chloride	2.84	84	5509	0.384 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100120\
Data File : VX018729.D
Acq On : 01 Oct 2020 17:46
Operator : JC/SP
Sample : VX1001WBL01
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK76

Quant Time: Oct 02 09:07:23 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR092820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 02 08:54:14 2020
Response via : Initial Calibration

