

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100121\
 Data File : VX024501.D
 Acq On : 01 Oct 2021 15:24
 Operator : JC/MD
 Sample : M3867-05
 Misc : 6.43g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-5-(1.5-2.5)

Quant Time: Oct 01 16:17:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	135340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	226396	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	222864	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.030	152	113180	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99069	51.938	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 103.880%			
35) Dibromofluoromethane	5.397	113	72992	48.532	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.060%			
50) Toluene-d8	8.653	98	279012	50.404	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.800%			
62) 4-Bromofluorobenzene	11.085	95	116060	53.325	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 106.640%			
Target Compounds						Qvalue
52) Toluene	8.720	92	6794	1.695	ug/l	95
67) Ethyl Benzene	10.195	91	14477	1.762	ug/l	99
68) m/p-Xylenes	10.305	106	28027	8.893	ug/l	96
69) o-Xylene	10.646	106	27113	8.700	ug/l	95
73) Isopropylbenzene	10.970	105	9152	1.144	ug/l	99
78) n-propylbenzene	11.305	91	37919	4.054	ug/l	98
80) 1,3,5-Trimethylbenzene	11.457	105	167981	24.575	ug/l	100
84) 1,2,4-Trimethylbenzene	11.756	105	439333	64.642	ug/l	99
85) sec-Butylbenzene	11.896	105	17534	2.080	ug/l	95
86) p-Isopropyltoluene	12.012	119	15055	2.094	ug/l	88
89) n-Butylbenzene	12.323	91	131686	20.687	ug/l #	11
95) Naphthalene	13.780	128	137059	17.351	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100121\
Data File : VX024501.D
Acq On : 01 Oct 2021 15:24
Operator : JC/MD
Sample : M3867-05
Misc : 6.43g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-5-(1.5-2.5)

Quant Time: Oct 01 16:17:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

