

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100121\
 Data File : VX024502.D
 Acq On : 01 Oct 2021 15:47
 Operator : JC/MD
 Sample : M3938-06ME
 Misc : 5.74g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SUB-SLAB-08(12-14)ME

Quant Time: Oct 01 17:25:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	152771	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	258896	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	251554	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129632	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107655	50.000	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.000%
35) Dibromofluoromethane	5.397	113	81458	47.362	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.720%
50) Toluene-d8	8.653	98	315670	49.867	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.740%
62) 4-Bromofluorobenzene	11.085	95	133455	53.620	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.240%
Target Compounds						Qvalue
64) Tetrachloroethene	9.275	164	3569	2.050	ug/l	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100121\
Data File : VX024502.D
Acq On : 01 Oct 2021 15:47
Operator : JC/MD
Sample : M3938-06ME
Misc : 5.74g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SUB-SLAB-08(12-14)ME

Quant Time: Oct 01 17:25:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

