

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100121\
 Data File : VX024506.D
 Acq On : 01 Oct 2021 17:21
 Operator : JC/MD
 Sample : M3975-09ME
 Misc : 13.78g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S-834-K2-SO-12-14-092821ME

Quant Time: Oct 02 04:08:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	145399	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	245018	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	227259	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105062	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	103234	50.377	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.760%
35) Dibromofluoromethane	5.397	113	77141	47.392	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.780%
50) Toluene-d8	8.653	98	296387	49.473	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.940%
62) 4-Bromofluorobenzene	11.085	95	112478	47.751	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.500%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.294	101	4244	2.956	ug/l	93
27) cis-1,2-Dichloroethene	4.489	96	69413	39.239	ug/l	91
44) Trichloroethene	7.129	130	121917	65.973	ug/l	100
64) Tetrachloroethene	9.281	164	1892	1.203	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100121\
Data File : VX024506.D
Acq On : 01 Oct 2021 17:21
Operator : JC/MD
Sample : M3975-09ME
Misc : 13.78g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S-834-K2-SO-12-14-092821ME

Quant Time: Oct 02 04:08:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

