

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100121\
 Data File : VX024507.D
 Acq On : 01 Oct 2021 17:45
 Operator : JC/MD
 Sample : M3975-10ME
 Misc : 10.34g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S-835-K2-SO-12-14-092821ME

Quant Time: Oct 02 04:09:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	149560	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	254582	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	243094	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117871	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	106495	50.523	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.040%			
35) Dibromofluoromethane	5.397	113	79866	47.223	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 94.440%			
50) Toluene-d8	8.653	98	308798	49.608	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 99.220%			
62) 4-Bromofluorobenzene	11.085	95	120788	49.353	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 98.700%			
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.300	101	1487	1.007	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	76057	41.799	ug/l	93
44) Trichloroethene	7.129	130	146164	76.122	ug/l	99
64) Tetrachloroethene	9.281	164	1793	1.066	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100121\
Data File : VX024507.D
Acq On : 01 Oct 2021 17:45
Operator : JC/MD
Sample : M3975-10ME
Misc : 10.34g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S-835-K2-SO-12-14-092821ME

Quant Time: Oct 02 04:09:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

