

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100124\
 Data File : VX043213.D
 Acq On : 01 Oct 2024 13:55
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 02 04:27:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 04:21:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	85958	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	161575	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	142883	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60367	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72453	48.969	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.940%
35) Dibromofluoromethane	5.385	113	54362	47.189	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.380%
50) Toluene-d8	8.647	98	197589	51.169	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.340%
62) 4-Bromofluorobenzene	11.079	95	71052	50.648	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.300%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100124\
Data File : VX043213.D
Acq On : 01 Oct 2024 13:55
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 02 04:27:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 04:21:52 2024
Response via : Initial Calibration

