

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
 Data File : VX047929.D
 Acq On : 01 Oct 2025 11:20
 Operator : JC/MD
 Sample : Q3193-06DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW4-20250922DL

Quant Time: Oct 03 01:17:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	124052	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	250181	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	254599	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	121354	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	96884	49.065	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.140%
35) Dibromofluoromethane	5.373	113	81226	48.411	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.820%
50) Toluene-d8	8.635	98	281463	48.480	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.960%
62) 4-Bromofluorobenzene	11.061	95	116365	52.812	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.620%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.117	130	25236	14.018	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
Data File : VX047929.D
Acq On : 01 Oct 2025 11:20
Operator : JC/MD
Sample : Q3193-06DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW4-20250922DL

Quant Time: Oct 03 01:17:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

