

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
 Data File : VX047932.D
 Acq On : 01 Oct 2025 12:22
 Operator : JC/MD
 Sample : Q3193-17DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D1-20250923DL

Quant Time: Oct 03 01:19:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	134999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	282843	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	293609	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	148894	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	114945	53.492	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.980%	
35) Dibromofluoromethane	5.379	113	94476	49.805	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.620%	
50) Toluene-d8	8.635	98	321946	49.050	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.100%	
62) 4-Bromofluorobenzene	11.061	95	138801	55.720	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 111.440%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.343	101	1288	0.825	ug/l	94
44) Trichloroethene	7.117	130	93082	45.733	ug/l	92
64) Tetrachloroethene	9.263	164	1118	0.584	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
Data File : VX047932.D
Acq On : 01 Oct 2025 12:22
Operator : JC/MD
Sample : Q3193-17DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D1-20250923DL

Quant Time: Oct 03 01:19:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

