

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
 Data File : VX047935.D
 Acq On : 01 Oct 2025 13:25
 Operator : JC/MD
 Sample : Q3193-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT20551-20250923

Quant Time: Oct 03 01:20:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	133993	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	279255	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	291356	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	141662	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	116327	54.541	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 109.080%			
35) Dibromofluoromethane	5.373	113	94137	50.264	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.520%			
50) Toluene-d8	8.635	98	321080	49.546	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 99.100%			
62) 4-Bromofluorobenzene	11.067	95	138083	56.144	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 112.280%			
Target Compounds						Qvalue
16) Acetone	2.380	43	1968	2.122	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	1698	0.292	ug/l	94
24) 1,1-Dichloroethane	3.611	63	1756	0.519	ug/l #	90
30) Chloroform	5.086	83	1403	0.410	ug/l	90
44) Trichloroethene	7.117	130	13307	6.622	ug/l	83
52) Toluene	8.708	92	5959	1.164	ug/l	93
64) Tetrachloroethene	9.263	164	595	0.313	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
Data File : VX047935.D
Acq On : 01 Oct 2025 13:25
Operator : JC/MD
Sample : Q3193-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT20551-20250923

Quant Time: Oct 03 01:20:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

