

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
 Data File : VX047936.D
 Acq On : 01 Oct 2025 13:45
 Operator : JC/MD
 Sample : Q3193-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB04-20250923

Quant Time: Oct 03 01:21:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	134914	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	272484	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	283583	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	146064	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	115870	53.956	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.920%	
35) Dibromofluoromethane	5.379	113	94315	51.611	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.220%	
50) Toluene-d8	8.635	98	313021	49.503	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.000%	
62) 4-Bromofluorobenzene	11.067	95	136111	56.718	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.440%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	22809	24.428	ug/l	94
25) 2-Butanone	4.544	43	61512	52.813	ug/l	97
29) Tetrahydrofuran	4.995	42	56298	80.349	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
Data File : VX047936.D
Acq On : 01 Oct 2025 13:45
Operator : JC/MD
Sample : Q3193-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB04-20250923

Quant Time: Oct 03 01:21:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

