

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
 Data File : VX047939.D
 Acq On : 01 Oct 2025 14:48
 Operator : JC/MD
 Sample : Q3193-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT162S1-20250924

Quant Time: Oct 03 01:23:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	119386	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	245473	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	248853	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	118778	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	96274	50.662	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.320%
35) Dibromofluoromethane	5.379	113	81734	49.648	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.300%
50) Toluene-d8	8.635	98	276805	48.593	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.180%
62) 4-Bromofluorobenzene	11.067	95	112879	52.213	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.420%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	1877	2.272	ug/l #	87
24) 1,1-Dichloroethane	3.605	63	4345	1.441	ug/l #	86
44) Trichloroethene	7.110	130	18483	10.464	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
Data File : VX047939.D
Acq On : 01 Oct 2025 14:48
Operator : JC/MD
Sample : Q3193-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT162S1-20250924

Quant Time: Oct 03 01:23:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

