

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
 Data File : VX047944.D
 Acq On : 01 Oct 2025 16:32
 Operator : JC/MD
 Sample : Q3194-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT150S1-20250922

Quant Time: Oct 03 01:25:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	126689	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	265155	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	277726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	141078	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	110113	54.604	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.200%	
35) Dibromofluoromethane	5.379	113	88961	50.027	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.060%	
50) Toluene-d8	8.635	98	299767	48.717	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.440%	
62) 4-Bromofluorobenzene	11.061	95	128652	55.091	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 110.180%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	1483	1.691	ug/l	90
30) Chloroform	5.080	83	1528	0.473	ug/l	93
44) Trichloroethene	7.117	130	19108	10.014	ug/l	93
64) Tetrachloroethene	9.257	164	1558	0.861	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
Data File : VX047944.D
Acq On : 01 Oct 2025 16:32
Operator : JC/MD
Sample : Q3194-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT150S1-20250922

Quant Time: Oct 03 01:25:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

