

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
 Data File : VX047946.D
 Acq On : 01 Oct 2025 17:14
 Operator : JC/MD
 Sample : Q3194-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D3-20250922

Quant Time: Oct 03 01:26:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	121038	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	247090	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	254262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	127057	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	99720	51.759	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.520%
35) Dibromofluoromethane	5.379	113	81231	49.019	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.040%
50) Toluene-d8	8.635	98	275232	48.000	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.000%
62) 4-Bromofluorobenzene	11.067	95	117101	53.811	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.620%
Target Compounds						
16) Acetone	2.380	43	1487	1.775	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100125\
Data File : VX047946.D
Acq On : 01 Oct 2025 17:14
Operator : JC/MD
Sample : Q3194-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D3-20250922

Quant Time: Oct 03 01:26:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

